



## Certificate of Analysis



### AzureSeq RUO qPCR Kit SARS-CoV-2 for 200 Reactions

Product name	AzureSeq RUO qPCR Kit SARS-CoV-2 for 200 Reactions
Reference number	AzureSeq-200 RUO
LOT number	OMX1866
Kit Assembling Date	2021.03.19.
Expiration date	2021.10.29
Below data are based on the following documents:	20R05006; 20U06011,20B30009,20B32008,20B31007,20B33007

### AzureSeq-200 RUO Components

Product Code	Product name	Pass/Fail*
OA-ITMP-MM-100	2X InhibiTaq Multiplex HotStart MasterMix	Pass
OA-RT-200	RTScript Reverse Transcriptase, 200U/uL	Pass
OA-CPPM-100uL	CoVi Primer/Probe Mix 3	Pass
OA-NFW-350uL	Nuclease Free Water	Pass
OA-CVNC-150	CoVi Negative Control	Pass
OA-CVPC-150	CoVi Positive Control	Pass

\*for acceptance criteria see second page

Authorization for release			
Name	T. Deák	Function	Laboratory Quality Coordinator
Signature		Sign date	2021.03.19.
Name	G. Adlovits	Function	RAQS Manager
Signature		Sign date	2021.03.19

TITLE: AZURESEQ-200 RUO\_v2

DOCUMENT No. COA-01- AZURESEQ-200 RUO

ISSUED: 02.10. 2021

STATUS: CURRENT

VERSION: 2

## Acceptance Criteria

Component	Assay name / Specification	Passing Result	Result
OA-ITMP-MM-100	<b>qPCR Amplification</b> 2X InhibitTaq Multiplex Master Mix is tested for performance in multiplex qPCR reactions using primers specific to the SARS-Cov2 N1/N2 EUA Assays, and RNase P with 100 copies of synthetic RNA and 1ng of UHRR. Successful amplification of synthetic RNA <40ct and UHRR <30ct	Corresponds	Corresponds
	<b>Unspecific dsDNA nucleases</b> HindIII digested Lambda DNA is incubated with 2x InhibitTaq Multiplex Master at a concentration of 1x at 37C for 16 hours. No detectable degradation of DNA via agarose gel electrophoresis.	Pass	Pass
	<b>N1/N2 EU amplicon Concentration:</b> 2X InhibitTaq Multiplex Master is used in multiplex qPCR reactions using primers specific to the SARA-Cov2 N1/N2 EUA Assay and RNase P. No amplification <40ct (<2%) is detectable.	Pass	Pass
OA-RT-200	<b>DNase and RNase Activity</b> 100U of RTScript RT is tested for nuclease degradation in reactions containing a DNA or RNA substrate. After incubation for 1 hour at 37C there is no detectable degradation of DNA or RNA substrate as determined by agarose gel electrophoresis.	Not Detected	Not Detected
	<b>Nicking Activity</b> pBR322 supercoiled plasmid DNA is incubated with 100 Units of RTScript at 37C for 4 hours. No detectable nicking activity as determined by agarose gel electrophoresis.	Not Detected	Not Detected
	<b>Activity</b> One unit is defined as the amount of enzyme required to catalyse the incorporation of 1 nmol of dTTP into an acid-insoluble form in 10 minutes at 37C. The test result is 200U/uL $\pm$ 15%	Corresponds	Corresponds
	<b>Performance Test</b> TScript Revers Transcriptase is tested for performance in a 20uL RT-qPCR assay. Ct values are within 1Ct of control enzyme for each serial dilution of template	Corresponds	Corresponds
	<b>Single Stranded Exonuclease Assay</b> ssDNS probes are incubated with 100U of RTScript Reverse Transcriptase for 4 hours at 37C.	Not Detected	Not Detected
	<b>Human gDNA Contamination Assay</b> 100U of RTScript Reverse Transcriptase is used as the template in qPCR reactions using primers specific to Human gDNA. No Human gDNA is detected before 45 cycles.	Pass	Pass
	<b>Shelf life</b>	$\geq$ 12 months	$\geq$ 12 months
OA-CPPM-100uL OA-NFW-350uL OA-CVNC-150 OA-COVPC-150	<b>qPCR amplification</b> Components are tested for performance in multiplex qPCR reaction using primers specific to the SARS-Cov2 N1/N2 EUA Assays, and RNase P with copies of synthetic RNA and 1ng of UHRR. Successful amplification os synthetic RNA <40ct and UHRR <30ct.	Pass	Pass
	<b>ssRNase</b> Single stranded RNA is incubated with the components at 37C for 2 hours. No degradation is observed.	Pass	Pass
	<b>Unspecific Endonucleases</b> dsEndonuclease: HindIII digested Lambda-DNA is incubated with the components 37C for 1 hour. No degradation is observed.	Pass	Pass
	<b>N1/N2 EU Amplicon concentration</b> Components are used in multiplex qPCR reactions using primers specific to the SARS-Cov2 N1/N2 EUA Assays and RNase P. No amplification <40ct (<2%) is detectable	Pass	Pass