



TITLE: CERTIFICATE OF ANALYSIS – OMIXON HLA TWIN RUO 3.1.0 - HOLOTYPE V1  
STATUS: CURRENT

DOCUMENT NO. COA-02

ISSUED: 2018-04-12

VERSION: 1

## Certificate of Analysis

Product name	Omixon HLA Twin RUO
Version number	3.1.0
Build ID	a7027383c76688e3be7a923eda0135819b7fc33c
Release date	6 <sup>th</sup> Sept, 2018
IMGT/HLA version	3.30.0_5
Assay version	Holotype HLA v1

### Test, acceptance criteria, result

Test	Test description	Acceptance criteria	Result
Genotyping result equivalency	Comparison of results generated with the tested software version compared to results generated with the previous public release.	Results are equal to or better than results generated with the previous version of the software (Omixon HLA Twin RUO 3.0.0).	PASSED
QC result equivalency	Comparison of QC results generated with the tested software version compared to results generated with the previous public release.	Results are equal to or better than results generated with the previous version of the software (Omixon HLA Twin RUO 3.0.0).	PASSED

Verified by			
Name	Krisztina Rigó	Function	Head of Software Research
Signature		Sign date	07/11/2018
Name	Gabriella Adlovits	Function	RAQS Manager
Signature		Sign date	07/11/2018