



TITLE: CERTIFICATE OF ANALYSIS – OMIXON HLA TWIN RUO 3.1.1 - HOLOTYPE V1
STATUS: CURRENT

DOCUMENT No. COA-02

ISSUED: 2018-04-12

VERSION: 1

Certificate of Analysis

Product name	Omixon HLA Twin RUO
Version number	3.1.1
Build ID	527911537bff7857ee376c96e33c5ea8da07db9c
Release date	22nd Oct, 2018
IMGT/HLA version	3.31.0_5
Assay version	Holotype HLA v1

Test, acceptance criteria, result

Test	Test description	Acceptance criteria	Result
Genotyping result equivalency	Comparison of results generated with the tested software version compared to results generated with the previous public release.	Results are equal to or better than results generated with the previous version of the software (Omixon HLA Twin RUO 3.1.0).	PASSED
QC result equivalency	Comparison of QC results generated with the tested software version compared to results generated with the previous public release.	Results are equal to or better than results generated with the previous version of the software (Omixon HLA Twin RUO 3.1.0).	PASSED

Verified by

Name	Krisztina Rigó	Function	Head of Software Research
Signature		Sign date	8 th Nov 2018
Name	Gabriella Adlovits	Function	RAQS Manager
Signature		Sign date	8 th Nov 2018